

Certificate

Producer inspection certificate M (Service Centre)

in accordance with DIN 55350-18-4.2.2

Product : TC307
Article no :
Serial no : 687283

Inspection date : 4.6.2002

Ordered by : Julius Pakozdy

Your order number : 200/6/2002

Customer : TSS Bratislava, a.s., Za stanickou 3, Bratislava

1. Specifications : Standard deviation of a mean direction
measured in both faces : 2 mGon

2. Test results : Standard deviation Hz : 0.38 mGon
Standard deviation V : 0.27 mGon

3. Certificate : We hereby certify that the product described has been tested and complies with the specifications and test results as stated above. The test equipment used is traceable to our working standards or to our working procedures. The working standards and the working procedures are traceable to national standards or to recognized procedures. This is established by our Quality Management System, audited to ISO 9001 by an independent national accredited body.

Geotech Bratislava
Cernyševského 26
851 01 Bratislava


4.6.2002




Julius Pakozdy
Manager Instr. Service & Documentation


Ivan Lovisek
Instr. Service & Documentation

Leica
Geosystems



Certificate

Producer inspection certificate M (Service Centre)

in accordance with DIN 55350-18-4.2.2

Product : TC307
Article no :
Serial no : 687283

Inspection date : 4.6.2002

Ordered by : Julius Pakozdy

Your order number : 200/6/2002

Customer : TSS Bratislava, a.s., Za stanickou 3, Bratislava

1. Specifications : Standard deviation of a mean direction
measured in both faces : 2 mGon

2. Test results : Standard deviation Hz : 0.38 mGon
Standard deviation V : 0.27 mGon

3. Certificate : We hereby certify that the product described has been tested and complies with the specifications and test results as stated above. The test equipment used is traceable to our working standards or to our working procedures. The working standards and the working procedures are traceable to national standards or to recognized procedures. This is established by our Quality Management System, audited to ISO 9001 by an independent national accredited body.

Geotech Bratislava
Cernyševského 26
851 01 Bratislava

4.6.2002




Julius Pakozdy
Manager Instr. Service & Documentation


Ivan Lovisek
Instr. Service & Documentation

Leica
Geosystems